

URGENT MEDICINE RECALL


EPIPEN 300 microgram (μg) Adrenaline Injection Syringe Auto-Injector Batches 5FA665, 5FA6651, 5FA6652, 5FA6653

Alphapharm Pty Ltd, following consultation with the Therapeutic Goods Administration is recalling the following batches of EpiPen® 300 microgram (μg) Adrenaline Injection Syringe Auto-Injectors (used for the treatment of allergic emergencies (anaphylaxis)) due to the potential that these devices may contain a defective part that may result in the device failing to activate or requiring increased force to activate. If the EpiPen fails to activate it may result in life threatening adverse events as the underlying anaphylaxis will not be treated.

| Batch number | Expiry |
|--------------|--------|
| 5FA665 | Apr 17 |
| 5FA6651 | Apr 17 |
| 5FA6652 | Apr 17 |
| 5FA6653 | Apr 17 |

IF YOU HAVE AN EPIPEN:

1. Check if you have a 300 μg EpiPen® (yellow carton and label) and if you do, check the batch number and expiry. The batch number and expiry can be found on the label of the pen or on the end of the carton.



2. If your EpiPen® 300 μg has the following batch numbers 5FA665, 5FA6651, 5FA6652 or 5FA6653 and an expiry of Apr 17, you need to replace it with a new one as soon as possible by returning to your pharmacist.
3. Your pharmacist will replace the EpiPen® 300 μg from the affected batch with an EpiPen® 300 μg from a different batch FREE OF CHARGE.
4. You must keep your current EpiPen® until you get a replacement and use it if required.

If your EpiPen® 300 μg is not from a batch listed above or is a green EpiPen® Jr 150 μg Adrenaline Injection Syringe Auto Injector, your product is not affected by this recall and no action is required.

Alphapharm Pty Ltd sincerely regrets any inconvenience to its customers