



# Subcutaneous immunoglobulin (SCIg) therapy

## Preparing the infusion set and inserting the Neria<sup>®</sup> Device (using Springfusor)

The Neria<sup>®</sup> device:



Flow control tubing for Springfusor syringe:



### 1. Preparing Springfusor syringe:

- a. Attach the flow control tubing to the tip of the Springfusor syringe. Check it is secure.




### 2. Preparing the Neria<sup>®</sup> device:

- a. Attach the tubing from the Neria device to end of the flow control tubing  
Take care not to touch the end connection of the tubing as you remove the Neria<sup>®</sup> device from the packet.



- b. To remove air bubbles and prime the line, keep the syringe upright and push the plunger of the syringe until the fluid fills the line.



If possible try to prime the line just up to the base of the needle. Keeping the needle “dry” may make needle insertion more comfortable.

### 3. Inserting the Needle

- Wipe away numbing (anaesthetic) cream from the skin at the injection site (if used).
- Wash and dry your hands.
- Hold the wings of the infusion set. Remove the adhesive backing and the protective cap.



- Gently pinch the skin and insert the needle at a 90 degrees angle. Ensure the needle is flush with the skin, and secure the adhesive dressing. Tape excess tubing in place, leaving some ‘give’ in case the tubing is pulled.



**PINCHING THE SKIN IS IMPORTANT TO ENSURE THE NEEDLE GOES INTO SUBCUTANEOUS (FATTY) TISSUE and NOT INTO MUSCLE**  
Remember always pinch an inch!

© ASCIA 2017

ASCIA is the peak professional body of clinical immunology and allergy specialists in Australia and New Zealand

#### Disclaimer

This document has been adapted with permission from resources developed by Princess Margaret Hospital Immunology Department (Department of Health, Western Australia). It has been peer reviewed by ASCIA members and is based on expert opinion and the available published literature at the time of review. Information contained in this document is not intended to replace medical advice and any questions regarding a medical diagnosis or treatment should be directed to a medical practitioner. Development of this document is not influenced by commercial organisations.